






23-JUL-2013 19:47:16	P80	Chlorine F&T PP	2.00 mg/L	
0.01 mg/L Cl ₂				
23-JUL-2013 19:46:45	P80	Chlorine F&T PP	2.00 mg/L	
-0.01 mg/L Cl ₂				
23-JUL-2013 19:46:01	P80	Chlorine F&T PP	2.00 mg/L	
0.01 mg/L Cl ₂				
23-JUL-2013 19:45:25	P80	Chlorine F&T PP	2.00 mg/L	
0.22 mg/L Cl ₂				
18-JUL-2013 23:10:31	Single λ			
0.078 Abs	83.6	%T	650 nm	
18-JUL-2013 23:10:21	Single λ			
0.078 Abs	83.6	%T	650 nm	
18-JUL-2013 23:10:03	Single λ			
0.086 Abs	82.0	%T	650 nm	
18-JUL-2013 23:06:15	Single λ			
0.000 Abs	100.0	%T	650 nm	
18-JUL-2013 23:05:59	Single λ	Absorbance > 3.5!		
***** Abs	*****	%T	650 nm	
18-JUL-2013 23:05:07	Single λ			
1.396 Abs	4.0	%T	650 nm	
18-JUL-2013 23:04:42	Single λ			
0.000 Abs	100.0	%T	650 nm	
18-JUL-2013 23:01:44	Single λ			
0.530 Abs	29.5	%T	650 nm	
03-JUL-2013 22:08:00	TNTplus 822 COD	1500 mg/L #12135		
***** mg/L COD		Unstable lighting conditions!		
03-JUL-2013 22:07:41	TNTplus 822 COD	1500 mg/L #12135		
***** mg/L COD		Unstable lighting conditions!		
03-JUL-2013 22:07:20	TNTplus 822 COD	1500 mg/L #12135		
***** mg/L COD		Unstable lighting conditions!		
18-JUN-2013 21:19:48	P645	Silica ULR	1000 µg/L	
4 µg/L SiO ₂				
10-JUN-2013 18:52:04	P645	Silica ULR	1000 µg/L	
0 µg/L SiO ₂				
10-JUN-2013 18:51:10	P645	Silica ULR	1000 µg/L	
0 µg/L SiO ₂				
28-MAY-2013 19:29:57	TNTplus 836 Nitrate	35.0 mg/L #12115		
0.485 mg/L NO ₃ ⁻ -N		Under measuring range!		
24-MAY-2013 22:54:31	TNTplus 822 COD	1500 mg/L #12135		
5673 mg/L COD		Over measuring range!		
17-MAY-2013 17:14:07	P9052	TIM		
-12.53 mg/L				
 Sample ID 2				
17-MAY-2013 17:13:41	P9052	TIM		
-12.54 mg/L				
 Sample ID 1				

10-APR-2013 21:04:36	P344	N Nitrate HR TNT	30.0 mg/L	
9.9 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:14:25	P344	N Nitrate HR TNT	30.0 mg/L	
11.1 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:14:16	P344	N Nitrate HR TNT	30.0 mg/L	
10.6 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:14:12	P344	N Nitrate HR TNT	30.0 mg/L	
47.1 mg/L NO ₃ ⁻				
10-APR-2013 20:11:58	P344	N Nitrate HR TNT	30.0 mg/L	
43.0 mg/L NO ₃ ⁻				
10-APR-2013 20:11:52	P344	N Nitrate HR TNT	30.0 mg/L	
9.7 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:11:38	P344	N Nitrate HR TNT	30.0 mg/L	
11.1 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:11:30	P344	N Nitrate HR TNT	30.0 mg/L	
10.6 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:11:10	P344	N Nitrate HR TNT	30.0 mg/L	
8.6 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:10:51	P344	N Nitrate HR TNT	30.0 mg/L	
38.3 mg/L NO ₃ ⁻				
10-APR-2013 20:10:35	P344	N Nitrate HR TNT	30.0 mg/L	
48.9 mg/L NO ₃ ⁻				
10-APR-2013 20:10:20	P344	N Nitrate HR TNT	30.0 mg/L	
46.8 mg/L NO ₃ ⁻				
10-APR-2013 20:09:43	P344	N Nitrate HR TNT	30.0 mg/L	
41.8 mg/L NO ₃ ⁻				
10-APR-2013 20:08:57	P344	N Nitrate HR TNT	30.0 mg/L	
9.4 mg/L NO ₃ ⁻ -N				
10-APR-2013 20:06:42	P344	N Nitrate HR TNT	30.0 mg/L	
11.5 mg/L NO ₃ ⁻ -N		Unstable lighting conditions!		
10-APR-2013 20:06:16	P344	N Nitrate HR TNT	30.0 mg/L	
50.7 mg/L NO ₃ ⁻		Unstable lighting conditions!		
10-APR-2013 20:02:21	P344	N Nitrate HR TNT	30.0 mg/L	
11.5 mg/L NO ₃ ⁻ -N		Unstable lighting conditions!		
10-APR-2013 19:40:51	P195	Fluoride AV	2.00 mg/L	
103.03 mg/L F				  100.00
10-APR-2013 19:39:49	P195	Fluoride AV	2.00 mg/L	
1.03 mg/L F				
10-APR-2013 19:04:05	P190	Fluoride	2.00 mg/L	
0.95 mg/L F				
10-APR-2013 18:54:14	P190	Fluoride	2.00 mg/L	
0.54 mg/L F				
10-APR-2013 18:48:11	P195	Fluoride AV	2.00 mg/L	
1.00 mg/L F				
10-APR-2013 18:38:26	P445	Oxygen Dis. HR AV	15.0 mg/L	
10.3 mg/L O ₂				

10-APR-2013 18:33:06	P445	Oxygen Dis. HR AV	15.0 mg/L	
9.7 mg/L O ₂				
10-APR-2013 18:17:28	Single λ			
0.618 Abs	24.1	%T	420 nm	
10-APR-2013 18:17:07	Single λ			
1.235 Abs	5.8	%T	420 nm	
10-APR-2013 18:13:16	P195	Fluoride AV	2.00 mg/L	
1.75 mg/L F ⁻				
10-APR-2013 18:11:51	P195	Fluoride AV	2.00 mg/L	
1.75 mg/L F ⁻				
10-APR-2013 18:11:00	P190	Fluoride	2.00 mg/L	
1.55 mg/L F ⁻				
10-APR-2013 18:07:51	P85	Chlorine F&T AV	2.00 mg/L	
1.68 mg/L Cl ₂				
10-APR-2013 18:07:39	P85	Chlorine F&T AV	2.00 mg/L	
0.86 mg/L Cl ₂				
10-APR-2013 18:07:24	P85	Chlorine F&T AV	2.00 mg/L	
0.20 mg/L Cl ₂				
09-APR-2013 22:02:55	P80	Chlorine F&T PP	2.00 mg/L	 Alex
1.28 mg/L Cl ₂				
09-APR-2013 22:02:32	P80	Chlorine F&T PP	2.00 mg/L	 Alex
1.29 mg/L Cl ₂				
09-APR-2013 21:58:17	P80	Chlorine F&T PP	2.00 mg/L	 Alex
1.24 mg/L Cl ₂				
09-APR-2013 21:57:43	P80	Chlorine F&T PP	2.00 mg/L	 Alex
1.24 mg/L Cl ₂				
09-APR-2013 21:56:55	P80	Chlorine F&T PP	2.00 mg/L	 Alex
1.26 mg/L Cl ₂				
09-APR-2013 20:17:28	P80	Chlorine F&T PP	2.00 mg/L	 Alex
6.18 mg/L Cl ₂		Over measuring range!		
01-APR-2013 18:34:10	TNTplus 836 Nitrate	35.0 mg/L #12115		
-0.215 mg/L NO ₃ ⁻ -N		Negative result!		
01-APR-2013 18:02:13	P80	Chlorine F&T PP	2.00 mg/L	
1.01 mg/L Cl ₂				
01-APR-2013 18:01:37	P80	Chlorine F&T PP	2.00 mg/L	
2.75 mg/L Cl ₂		Over measuring range!		
29-MAR-2013 23:36:52	P80	Chlorine F&T PP	2.00 mg/L	
nan mg/L Cl ₂ SA				
18-MAR-2013 20:01:46	P190	Fluoride	2.00 mg/L	
-2.07 mg/L F ⁻		Under measuring range!		
07-MAR-2013 18:45:18	P80	Chlorine F&T PP	2.00 mg/L	
0.00 mg/L Cl ₂				
 WOODY				
07-MAR-2013 18:43:52	P80	Chlorine F&T PP	2.00 mg/L	
0.00 mg/L Cl ₂				
01-MAR-2013 21:20:02	P88	Chlorine F&T HR	10.0 mg/L	