

UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

AUG 19 2003

Cary B. Jackson, Ph.D.  
Regulatory Chemist  
Hach Company  
5600 Lindburgh Drive  
PO Box 389  
Loveland, Colorado 80539-0389

Dear Dr. Jackson:

Per your request, this letter confirms that extension of the calibration range of an EPA approved or accepted test procedure (analytical method) is within the scope of the method and does not require regulatory approval, provided that the upper limit of the linear range of the instrument or analytical system is not exceeded.

Sincerely yours,

*for* *Marion Kelly*  
William A. Telliard, Director  
Analytical Methods  
Office of Science and Technology (MC 4303T)

CONCURRENCES							
IBOL	SAB						
SURNAME	Kelly						
DATE	8/19/03						

FDA Form 1027-1A (Rev. 11/97) OFFICIAL FILE COPY  
TOTAL P.01